

# Test Report

## – Eppendorf PCR clean & Protein free –

Blank example certificate

• **Human DNA**

Method: *real-time* PCR acc. to SOP 99880/V8  
Limit: < 2 pg; less than one human cell

• **DNase**

Method: DNA Digestion acc. to SOP 100570/V4  
Limit: Not detectable  
(LOD:  $1.0 \times 10^{-7}$  Kunitz units)

• **RNase**

Method: RNA Digestion acc. to SOP 100571/V4  
Limit: Not detectable  
(LOD:  $1.0 \times 10^{-9}$  Kunitz units)

• **PCR Inhibition**

Method: *real-time* PCR acc. to SOP 100559/V4  
Limit: Less than 10 targets amplifiable

• **Protein**

Method: Bradford-Test (Ph. Eur. 10.0 / 2.5.33)  
Limit: Not detectable  
(LOD: 1 µg/mL)

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination:

Testing of the rinse solution of the above described lot showed conformity within the limits of detection.

\_\_\_\_\_  
Date of release

\_\_\_\_\_  
Registration number

\_\_\_\_\_  
Dr. rer. nat. Sybille Kuhnert (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

page 1 of 1