



UK Health
Security
Agency

Certificate of Containment Testing

Containment Testing of Rotor R12A3 in a High-Speed Centrifuge

Report No. 23/013E

Report Prepared For: Eppendorf SE, Hamburg, Germany

Issue Date: 28 November 2023

Test Summary

Rotor R12A3 (max. speed 12 000 rpm) was containment tested at 12 000 rpm in an Eppendorf Himac Technologies High-Speed centrifuge CR22N, using Annex AA of IEC 61010-2-020:2016 (3rd Ed.). The sealed rotor was shown to contain a spill within the rotor.

Report Written By

Name: Ms Helen Hookway

Title: Biosafety Team Leader

Report Authorised By

Name: Mr Thomas Pottage

Title: Senior Biosafety Scientist