

Declaration of Conformity

The product named below fulfills the requirements of directives and standards listed. In the case of unauthorized modifications to the product or an unintended use this declaration becomes invalid. This declaration of conformity is issued under the sole responsibility of the manufacturer.

Product name:

CellXpert® C170i, CellXpert® C170

including components

Product type:

CO₂-Incubator

Relevant directives / standards:

2014/35/EU: EN 61010-1 +A1 + A1/AC, EN IEC 61010-2-010, EN 62233, EN IEC 62311

2014/30/EU: EN 61326-1


2011/65/EU: EN IEC 63000
(incl. (EU) 2015/863)

Further applied standards: IEC 61010-1 + AMD1, IEC 61010-2-010, IEC 62233, IEC 62311
UL 61010-1, UL 61010-2-010
CAN/CSA C22.2 No. 61010-1-12, CAN/CSA C22.2 No. 61010-2-010-19
IEC 61326-1

Hamburg, June 08, 2022



Dr. Wilhelm Plüster
Management Board



Dr. Marlene Jentzsch
Senior Vice President
Division Separation & Instrumentation

Your local distributor: www.eppendorf.com/contact
Eppendorf SE · Barkhausenweg 1 · 22339 Hamburg · Germany
eppendorf@eppendorf.com

ISO
9001
Certified

ISO 13485
Certified

ISO 14001
Certified

Eppendorf®, the Eppendorf Brand Design and CellXpert® are registered trademarks of Eppendorf SE, Germany.
All rights reserved, incl. graphics and pictures. Copyright ©2022 by Eppendorf SE.